



Technical Bulletin

178 Blackstone Avenue Jamestown NY 14701

www.jamestown.com

716.665.5313

Date: *September 1, 2011* No.: 028
To: Customer Service, Engineering, Estimating, Operations, Sales, Valued JMP Dealers
Subject: *Fume Hoods with Smart-Sash™, Impact on UL 1805 Certification*

Please be advised that the addition of the JMP Smart-Sash™ system (or equivalent automatic sash control system) to JMP produced and UL 1805 listed Fume Hoods, if fully installed and operational, voids our Manufacturers UL 1805 certification for those hoods and prohibits us from applying the UL logo.

JMP will continue to apply the UL logo to listed Fume Hoods. However, it will be necessary for the field installer to complete the installation and make the JMP Smart-Sash™ system (or equivalent automatic sash control system) operational per the instructions provided with the Fume Hoods.

We are in the process of adding the JMP Smart-Sash™ system to our UL 1805 listing and/or getting UL approval of the sub-system as an add-on to our UL 1805 listed Fume Hoods. We will keep you advised of our progress and forecasted time line to completion as these details become available.

Contact JMP for more information.

*Fume Hood Engineering
Jamestown Metal Products, LLC*